



BHARAT COKING COAL LIMITED
(A Subsidiary of Coal India Ltd.)

SR No:-124&125

भारत कोकिंग कोल लिमिटेड
OFFICE OF CHAIRMAN-CUM-MANAGING-DIRECTOR
Bharat Coking Coal Limited,,



P.O - BCCL Township, Dist.-Dhanbad, Jharkhand, India-826005

Coal India Limited

A Maharatna Company
www.coalindia.in

Phone No:

Fax :

Website :

CIN : U10101JH1972GOI000918

Email :

GSTN : 20AAACB7934MFZB

Purchase Order No: 5100001072

PURCHASE ORDER

Document Date: Nov 26, 2021

To,

MEDICHEM ELECTRONICS

MEDICHEM ELECTRONIC SB-60 ELECTRONIC
CITY, NOIDA GAUTAM BUDDHA NAGAR

201301

Uttar Pradesh

India

Department: MM

PCC BY REGD. POST/MAIL

E-Mail Address : MEDICHEMELECTRONICS@GMAIL.COM

Telephone :

GST No : 09AFPPB3580C1ZK

Vendor Code : 10001167

Purchase Order for procurement of Pediatric & Neonatal Equipments for CHD, BCCL.

As per GeM Contract Ref no.

1. GEMC-511687782381404 dated 26.11.2021 for Clinical incubators or infant warmers (Q3) - 07 Nos.
2. GEMC-511687721415548 dated 26.11.2021 for Phototherapy machine for SNCU (Q3) - 03 Nos.

Item No	Material No.	Short Text	HSN/SAC	UOM	Quantity	Unit Price	Price (Rs)
10		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
20		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
30		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							

40		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
50		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
60		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
70		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
80		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
90		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
100		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
						Total Net Value Exclusive of Tax	INR 352850.00
						Total Net Value Inclusive of Tax	INR 352850.00
Amount in words:	Three Lakh Fifty Two Thousand Eight Hundred Fifty Rupees Only						

General Terms and Conditions

As per GeM Contract Ref no.

1. GEMC-511687782381404 dated 26.11.2021 for Clinical incubators or infant warmers (Q3) - 07 Nos.
2. GEMC-511687721415548 dated 26.11.2021 for Phototherapy machine for SNCU (Q3) - 03 Nos.

Contract



Contract No: GEMC-511687782381404

Generated Date: 26-Nov-2021

Bid/RA/PR No: [GEM/2021/B/1475313](#)

Schedule Name: Schedule 1

Organisation Details	Buyer Details
Type: Central PSU Ministry: Ministry of Coal Department: Materials Management Organisation Name: Bharat Coking Coal Limited Office Zone: Bccl Dhanbad	Designation: Assistant Manager6 Contact No.: 62876-95241- Email ID: mayank.shekhar@nic.in GSTIN: 20AAACB7934MFZB Address: MM Department, level 3, Commercial Building, Koyla Bhawan, Koyla Nagar, DHANBAD, JHARKHAND-826005, India

Financial Approval Detail	Paying Authority Details
IFD Concurrence: Yes Designation of Administrative Approval: D (T) P&P Designation of Financial Approval: GM (MM) / HOD	Payment Mode: Offline Designation: HOD Purchase Finance Email ID: pao5.bccl.jh@gembuyer.in GSTIN: N Address: MM Department, level 3, Commercial Building, Koyla Bhawan, Koyla Nagar, DHANBAD, JHARKHAND-826005, India

Seller Details	
GeM Seller ID: 90AC180000614810 Company Name: MEDICHEM ELECTRONICS Contact No.: 09811139456 Email ID: medichemelectronics@gmail.com Address: B-60, Sec-64, Electronic City, Noida, Gautam Budh Nagar, Ghaziabad, UTTAR PRADESH-201301, - MSME verified: Yes MSME Registration number: UP28B0014569 MSE Social Category: General MSE Gender: Male GSTIN: 09AFPPB3580C1ZK , 09AFPPB3580C1ZK , 09AFPPB3580C1ZK	

*GST / Tax invoice to be raised in the name of - Consignee

Product Details								
#	Item Description	Category Name & Quadrant	Model	HSN Code	Ordered Quantity	Unit	Lead Time(Days)	Price (Inclusive of all Duties and Taxes in INR)
1	Product Name : MEDILAP quartz infrared Clinical infant warmers Brand : MEDILAP Brand Type : Registered Brand Catalogue Status : OEM verified catalogue Selling As : OEM	Clinical incubators or infant warmers (Q3)	MRW-02+	HSN not specified by seller	7	pieces	-	272,300
Total Order Value (in INR)								272,300

Consignee Detail						
S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Designation: - Email ID: anand.kumar@coalindia.in Contact: 0326-2206167-234 GSTIN: 20AAACB7934MFZB Address: Chief Of Medical Services, Central Hospital , PO- Jagjivan Nagar., DHANBAD, JHARKHAND-826003, India	MEDILAP quartz infrared Clinical infant warmers	-	7	26-Nov-2021	10-Jan-2022

Product Specification for MEDILAP quartz infrared Clinical infant warmers		
Specification	Sub-Spec	Value



BHARAT COKING COAL LIMITED
(A Subsidiary of Coal India Ltd.)

SR No:-124&125

भारत कोकिंग कोल लिमिटेड
OFFICE OF CHAIRMAN-CUM-MANAGING-DIRECTOR
Bharat Coking Coal Limited,,



P.O - BCCL Township, Dist.-Dhanbad, Jharkhand, India-826005

Coal India Limited

A Maharatna Company
www.coalindia.in

Phone No:

Fax :

Website :

CIN : U10101JH1972GOI000918

Email :

GSTN : 20AAACB7934MFZB

Purchase Order No: 5100001072

PURCHASE ORDER

Document Date: Nov 26, 2021

To,

MEDICHEM ELECTRONICS

MEDICHEM ELECTRONIC SB-60 ELECTRONIC
CITY, NOIDA GAUTAM BUDDHA NAGAR

201301

Uttar Pradesh

India

Department: MM

PCC BY REGD. POST/MAIL

E-Mail Address : MEDICHEMELECTRONICS@GMAIL.COM

Telephone :

GST No : 09AFPPB3580C1ZK

Vendor Code : 10001167

Purchase Order for procurement of Pediatric & Neonatal Equipments for CHD, BCCL.

As per GeM Contract Ref no.

1. GEMC-511687782381404 dated 26.11.2021 for Clinical incubators or infant warmers (Q3) - 07 Nos.
2. GEMC-511687721415548 dated 26.11.2021 for Phototherapy machine for SNCU (Q3) - 03 Nos.

Item No	Material No.	Short Text	HSN/SAC	UOM	Quantity	Unit Price	Price (Rs)
10		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
20		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
30		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							

40		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
50		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
60		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
70		Clinical incubators or infant warmers		EA	1.000	38900.00 INR	38900.00
						Gross Price	38900.00
Material Description :							
Part No. :							
80		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
90		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
100		Phototherapy machine for SNCU		EA	1.000	26850.00 INR	26850.00
						Gross Price	26850.00
Material Description :							
Part No. :							
						Total Net Value Exclusive of Tax	INR 352850.00
						Total Net Value Inclusive of Tax	INR 352850.00
Amount in words:	Three Lakh Fifty Two Thousand Eight Hundred Fifty Rupees Only						

General Terms and Conditions

As per GeM Contract Ref no.

1. GEMC-511687782381404 dated 26.11.2021 for Clinical incubators or infant warmers (Q3) - 07 Nos.
2. GEMC-511687721415548 dated 26.11.2021 for Phototherapy machine for SNCU (Q3) - 03 Nos.

Technical Characteristics (specific to this type of device)

facility to display skin set, skin observed temperature in degree C and heat power separately	Yes
have user friendly touch panel control	Yes
Type of heater:	quartz infrared
Audiovisual alarm facility for:	Overheating beyond set temperature range,Spurt in Heater output i.e. more than 60% heater output for 10 minutes system should raise alarm and cutoff heater,Power failure,Heater failure,Patient temperature less/Greater than required/set temperature,Probe failure
Warmer head should be rotatable in different direction, so as to allow taking X-ray	Yes
Observation light Luminance and Colour temperature	> = 1000 Lux , Colour temperature range from 3700K to 5100K
Battery back up facility for Power failure indication during power fail	Yes
desired temperature range	25 to 40 degree C
setable temperature 32 to 38 deg C	32 to 38 deg C
temperature resolution should be:	<= 0.1 degree C
Temperature accuracy should be	<= 0.2 degree C
facility to lock the keyboard to avoid unwanted user modification of the set parameters	Yes
should have separate Bassinet trolley, bed should be tilt-able and have suitable provision for x-ray cassette holder, Mattress foam density should be minimum 25 kg/cm ³ , transparent collapsible side walls easily detachable for cleaning	Yes
Should have a Feather Touch operation with large digital display and comprehensive alarms, Control Panel should be liquid proof and allow easy and hygienic disinfection	Yes
Manual Mode can adjust Heater Output 10 -100 %, with 10% increment, an auditory and visual alarm shall be given at least every 15 min	Yes
Under manual mode, heater cut off / switch off , if the maximum irradiance at any point of the mattress area exceeds a total irradiance level of 10 mW/cm ² (between 10 to 30 minutes)	Yes
Bed height from Floor	90
Bed Distance from the heat source	90
should have lockable castor wheels	Yes
Green indicator light shall be indicate that warmer is ready for normal use	Yes
Markings on the bassinet and X-Ray cassette holder to enable proper positioning of the baby while doing the X-Ray	Yes
Number of tubing ports (edges covered by silicon rings) on the side walls (For cable management)	4
Height of the side walls over the mattress	110
Provision of X-Ray cassette tray of at least 750 mm X 350 mm size, suitable to adopt <= 20mm thick X-Ray cassette	Yes
bay bed should be crevice free for ease of cleaning, infection control	Yes

	mattress used should be of biocompatible material	Yes
	Skin temperature probe should be small in size [Diameter <= 10 mm & Height <= 4 mm] with biocompatible Baby contact material	Yes
	Mobility, portability	On castors (2 of the castors should have brakes) & Castor size >= 4 inch.
Definition and Clinical Purpose	Definition	Mains electricity (AC-powered) mobile device that contains an infrared (IR) heating element(s) designed to emit controlled, evenly distributed overhead heat to the body of a newborn/infant patient requiring supplemental heat. This device is equipped with wheels so that it can easily be moved to different areas of a room, ward, or department.
	Clinical purpose & Overview of functional requirements	It is an electrically powered device with a radiant heating source intended to maintain the thermal balance of an infant by direct radiation of energy in the infrared region of the electromagnetic spectrum. It is a microprocessor controlled unit with heater placed on the over head panel. This work on both servo and manual mode options to maintain the baby temperature at the set value. There are two modes of operation manual and baby (servo) mode. It has Digital displays reading the set and baby observed temperatures separately.
Settings	Settings with facility to clearly display the selected mode, Option of Selecting either Manual mode and Baby (Servo) mode settings, Provision for Set temperature range (in servo mode) from 32 to 38 deg C	Yes
	User's interface	Manual and Servo controlled temperature regulation
	Software and/or standard of communication (where ever required)	LED Display and inbuilt software; Interruption and restoration of the power supply does not change the preset values
	Others:	1. Device shall not overbalance when placed in any transport position of normal use on a 10° inclined plane from the horizontal plane. 2. Transformers of devcie shall be protected against overheating in the event of short circuit or overload of any output winding. 3. Patient leakage current should be less than 100 µA in normal condition 4. Temperature on the baby mattress should not exceed 43 deg C when the warmer is operating under steady temperature condition 5. Temperature of HEATER GUARDS should not exceed 85 °C in normal use. 6. The Temperature differences on the mattress shall not exceed 2 °C.
PHYSICAL CHARACTERISTICS	Overall Height of equipment	1800
	Overall Width of equipment	450
	Overall Length of equipment	600
	Overall weight of equipment	50
	Heat source Configuration	>= 60 degree angle adjustment must be possible in the heat source, it should provide shielding to the infant in case of breakage of tubes/bulbs and all surfaces to be made of corrosion resistant material
	Noise	Auditory alarm shall have a sound level of at least 65 dBA at a distance of 3 m from the front of the infant radiant warmer, and the sound level of the alarm shall not exceed 80 dBA on the mattress
	Heat dissipation	Should maintain desired set temperature and uniform heat distribution in baby cradle
ENERGY SOURCE	Power Requirements	220 to 240V, 50 Hz AC, ± 10% of input
	Battery operated	Power failure indication during power fail
	Protection	OVP, earth leakage protection
	Power consumption	650
	Compatible to alternate energy supply source; Solar Heating in addition to AC power source):	Yes
	Thermal reflector to fix the skin probe	>= 50 units to be supplied with each equipment unit
ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS	Atmosphere / Ambiance	Operating condition: -Capable of operating continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances. - an ambient air velocity is less than 0.3 m/s.
	User's care, Cleaning, Disinfection & Sterility issues	Complete unit to be easily washable and sterilizable using both alcohol and chlorine agents.
STANDARDS AND SAFETY	Certificates,	CE European
	Performance and safety standards (specific to the device type); Local and/or international	IEC-60601-1-2:2007 Medical electrical equipment Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility -Requirements and tests (Or Equivalent BIS). Shall meet IEC 60601 -2- 21: 2009 Medical Electrical Equipment - Part 2-21: Particular Requirement for the basic safety and essential performance of infant radiant warmers . should meet IEC 60601-1:2005(or latest) standard requirements. Baby contact material should be biocompatible. Manufacturer should be ISO 13485 certified
	Availability of Type Test report from any Govt/ NABL accredited /ILAC approved laboratory (To be produced by Seller on Buyer's demand if	Yes

	indicated available)	
TRAINING AND INSTALLATION	Pre-installation requirements for the equipment	Availability of 5 amp/15 Amp. Electrical socket (2 nos) for each warmer at Buyer's facility.
	Documentation required to be provided by Seller:	User & Operating manual, Certificate of Calibration, Factory inspection report, user training manual
WARRANTY	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	3

Terms and Conditions

1. Special terms and conditions- Version:1 effective from 03-09-2020 to 07-09-2021

1.1 1. Comprehensive warranty:

Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.

2. Service centres:

Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

3. Source of supply:

It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them .

4. Packing and Marking: Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.

5. Spare Parts:

Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied

6. Installation, Training, Manuals:

Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable.

7. Electrical safety checking:

Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.

8. Software:

All software updates should be provided free of cost during warranty period.

2. General Terms and Conditions-

2.1 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) as available on the GeM portal (unless otherwise superseded by Goods / Services specific Special

Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

2.2 Terms of delivery: Free Delivery at Site including loading/unloading. In respect of items requiring installation and / or commissioning and other services in the scope of supply (as indicated in respective product category specification / STC / ATC), and the cost of the same is also included in the Contract price.

2.2.1 Contracted goods should be delivered at the consignee or designated delivery location as per the working time of the buying organisation. Seller may get the same confirmed from consignee before scheduling delivery.

2.2.2 A copy of the contract should be available with the messenger / dispatching agency that delivers the Goods at consignee / delivery location (preferably pasted / attached outside the consignment / package) for easy reference and ease in delivery acceptance.

2.3 Delivery period: The Delivery Period/Time shall be essence of the Contract and delivery must be completed not later than such date(s). Any modification thereto shall be mutually agreed and incorporated in the Contract as per the provisions of the GTC.

2.4 Performance Security: If the Seller fails or neglects to observe or perform any of his obligations under the contract it shall be lawful for the Buyer to forfeit either in whole or in part, the Performance Security furnished by the Seller.

2.5 Taxes and Duties: Contract Prices are all inclusive i.e. including all taxes, duties, local levies / transportation / loading-unloading charges etc. Break up of GST shall be indicated by the Seller while raising invoice / bill on GeM. While submitting the bill / invoice Seller shall undertake that the Goods and Services Tax (GST) charged on this bill is not more than what is payable under the provision on the relevant Act or the Rules made there under and that the Goods on which GST has been charged have not been exempted under the GST Act or the Rules made there under and the charges on account of GST on these goods are correct under the provision of that Act or the rules made there under.

2.6 Octroi Duty and / or other local taxes:Contract Prices are all inclusive hence no reimbursement over and above the contract price(s) shall be allowed to seller towards payment of local taxes (such as levy of town duty, Octroi Duty, Terminal Tax and other levies of local bodies etc).

2.7 Limitation of Liability:The provisions of limitation of liability between Buyer and Seller as given in the GTC shall be applicable here.

2.8 Resolution of disputes: The provisions of DISPUTE RESOLUTION BETWEEN BUYER AND SELLER as given in the GTC shall be applicable here.

2.9 Liquidated Damages: If the Seller fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% per week or part of the week of delayed period as pre-estimated damages not exceeding 10% of the contract value without any controversy/dispute of any sort whatsoever. In case, Service Level Agreement (SLA) is applicable the same shall be applicable for the Contract.

2.10 Financial Certificate:

2.10.1 The expenditure involved for this purpose has received the Sanction of the competent financial authority.

2.10.2 The funds are available under the proper head in the sanction budget allotment for the concern financial year.

2.10.3 I have been fully authorized by the department to sign the supply order or incur the liability of the Goods being ordered.

2.11 The bidder should submit a self declaration to the effect in bidder's official letter head that their agency have not been black listed by any Agency whatsoever till date.

3. Buyer Added Bid Specific Terms and Conditions-

3.1 Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

3.2 Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 25% of total value.

3.3 Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

3.4 The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

3.5 Warranty period of the supplied products shall be as given in specifications from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

3.6 Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.