

**Organisation Details**

Type: Central PSU  
 Ministry: Ministry of Coal  
 Department: Materials Management  
 Organisation Name: Bharat Coking Coal Limited  
 Office Zone: Bcl Dhanbad

**Buyer Details**

Name: Ashok Kumar Pandey  
 Designation: Senior Manager(MM)  
 Contact No.: -  
 Email ID: ak.pandey4576@nic.in  
 GSTIN: N  
 Address: MM Department, level 3, Commercial Building, Koyla Bhawan, Koyla Nagar, DHANBAD, JHARKHAND-826005, India

**Financial Approval Detail**

IFD Concurrence : No  
 Designation of Administrative Approval : D(T) PP, BCCL  
 Designation of Financial Approval : D(T)PP, BCCL

**Seller Details**

Company Name: GOLDEN STAR SURGICAL INDUSTRIES PRIVATE LIMITED  
 Contact No.: 09888632327  
 Email ID: san66471@yahoo.co.in  
 Address: Plot No. 2, Sanghal Road, Adjoining Leather Complex, Jalandhar, Punjab-144002, -  
 MSME verified: Yes  
 MSE Social Category: General  
 MSE Gender: Male  
 GSTIN: 03AAGCG6037N1ZG

**Product Details**

#	Item Description	Category Name	Model	HSN Code	Ordered Quantity	Unit	Price (Inclusive of all Duties and Taxes in INR)
1	GSSI Electro-Mechanical (Motorized ) Delivery Bed	Delivery Bed	LABOUR DELIVERY BED	9402	1	pieces	68,500
Total Order Value (in INR)							68,500

**Consignee Detail**

S.No	Consignee	Item	Quantity	Expected Delivery Date
1	Rajendra Prasad - Landline-0326-2206167-234 rajendra.p2183@nic.in Chief Of Medical Services, Central Hospital , PO- Jagjivan Nagar., DHANBAD, JHARKHAND-826003, India	GSSI Electro-Mechanical (Motorized ) Delivery Bed	1	27-Apr-2020

**Product Specification for GSSI Electro-Mechanical (Motorized ) Delivery Bed**

Specification	Sub-Spec	Value
Standards	Conformity to standard for product	European CE
	Certification number and date	CE-BV-1904-1378,23-04-2019
	Conformity to standard for Electromagnetic compatibility & safety	IEC-60601-1-2:2001 Standard /IS 13450
	Purpose	It used in Obstetrics/Gynecology department for baby delivery in labour room with comfortable requirement of patient and easy accessible professional need for delivery team

Performance Parameters

Bed movement functionality	Control with remote
Type of delivery bed actuator	Electro-Mechanical ( Motorized )
Type of mechanism for controlling angular motion of bed	Hydraulic mechanism
Type of mechanism for functioning or controlling height of bed	Hydraulic mechanism
Number of section provided in bed top	3
Large perineal cut in C shape must be provided in seat section	Yes
Side safety belt provided	Yes
Type of side panel	Collapsible type
Number of Side panel	2
Head panel/Board provided	Yes
Shape of Head panel	C-type
Type of Head Panel	Detachable type
Foot panel provided	Yes
Shape of Foot panel	C-type
Type of foot Panel	Detachable type
Provision to convert bed in to table	Yes
Good grip provided to head/foot or side panel	Yes
Delivery bed should be Hand grip /push grip handles on both side	Yes
Adjustable leg rest provided	Yes
Lithotomy position & padded knee crutches provided	Yes
Slots for IV rod a head section provided	Yes
Chromium plated IV rod provided	Yes
Number of hooks provided in IV Rod	2
Power of motor in HP	0.5 hp
Availability of Noiseless, non rusting swivel castor wheel/ roller with for easy Mobility & Steering	Yes
Braking and locking mechanism provided in castor roller	Yes
No of castor to which braking system provided	2
Corner buffer provided in all four corner to protect patient from collision shock	Yes
Functionality of Foot side panel	Detachable type
Safe working load capacity of delivery bed in kg	150
Facility of instant non powered/mechanical CPR( Cardiopulmonary Resuscitation) release at the head end for emergency	NA
Urine Bag holder provided	Yes
Mattress provided with delivery bed	Yes
Mattress foam of the Bed should be with Anti Microbial agent incorporated into all components that assists in Prohibiting growth of bacteria & fungi and easy to clean	Yes
Mattress shall be made of High resilient & bio-density foam	Yes
Operating temperature & Humidity	10 -40 Degree Celsius and relative humidity of 15-90%
"Power Supply "	Single phase 220-240 Volt, 50 Hz ( AC Supply )
Spares Pair of Bed Ends, detachable supplied	Not supplied

	Spares Mattress	Not supplied
	Facility of Resettable overcurrent breaker shall be fitted for protection	Yes
	Electric Shock Protection level	Class-B/Class -1
	Warranty in years	3
	Base frame and support frame should be fabricated using metallic square / rectangular section of adequate cross section and thickness to provide high structural strength and stability	Yes
	Delivery Bed should be made up of high quality metal, pretreated with materials providing good finish, scratch resistant, bacteriostatic coating	Yes
	Type of powder/Paint coating done to delivery Bed frame,top & side panel	Epoxy Polyester Coating 50-60 microns (In Case of MS)
	Process used for pre chemically treatment of metallic bed component	Dip tank processes (In Case of MS)
	Provision of easily cleaning /sterilize(Especially Blood Stains)/maintenance to delivery Bed	Yes
	Product instruction manual shall be supplied	Yes
	Demonstration of equipment and training to be provided after completing supplies before Acceptance	Yes
Material Parameters	Material for the frame of bed	MS
	Material Side Safety belt	Velcro
	Material for Head & Foot Panel/Board	ABS Plastic
	Material for Bed top Section	Acrylic Sheet
	Material of wheels	SS
	Material for IV Rod	SS
	Material for bowl	SS
	Material for waste receptacle bucket	SS
	Material for delivery Bed Mattress	PU Foam
Dimensional Parameters	Maximum Adjustable Back Rest Angle in Degree	60
	Maximum Trendlenburg Angle in Degree	30
	Maximum Reverse Trendlenburg Angle in Degree	30
	Adjustable bed height in mm with 5% tolerance	550-850
	Diameter of Castor Roller in mm	100
	Clearance between Bed Base frame and Floor surface in mm	125
	Maximum thickness of Mattress in mm with tolerance of plus minus 2mm	74
	Approx foam Density in kg/m3 tolerance	30-50
	Therapeutic Weight limit for mattress in kg	125
	Length range of delivery bed in mm	2001-2150
	Width range of delivery bed in mm	800-900
	Weight of delivery Bed in Kg	140
Reports	Availability of ISO certification for manufacturer	Yes
	ISO certification No and Date if ISO certificate not available then put NA	M-RS-1704-1857
	Copies of certificates and other certifications to be furnished to buyer on demand at time of supplies	Yes

**Corrigendum**

## Terms and Conditions

---

### 1. General Terms and Conditions

**1.1** This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) as available on the GeM portal (unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

**1.2 Terms of delivery:** Free Delivery at Site including loading/unloading. In respect of items requiring installation and / or commissioning and other services in the scope of supply (as indicated in respective product category specification / STC / ATC), and the cost of the same is also included in the Contract price.

**1.3 Delivery period:** The Delivery Period/Time shall be essence of the Contract and delivery must be completed not later than such date(s). Any modification thereto shall be mutually agreed and incorporated in the Contract as per the provisions of the GTC.

**1.4 Performance Security:** If the Seller fails or neglects to observe or perform any of his obligations under the contract it shall be lawful for the Buyer to forfeit either in whole or in part, the Performance Security furnished by the Seller.

**1.5 Taxes and Duties:** Contract Prices are all inclusive i.e. including all taxes, duties, local levies / transportation / loading-unloading charges etc. Break up of GST shall be indicated by the Seller while raising invoice / bill on GeM. While submitting the bill / invoice Seller shall undertake that the Goods and Services Tax (GST) charged on this bill is not more than what is payable under the provision on the relevant Act or the Rules made there under and that the Goods on which GST has been charged have not been exempted under the GST Act or the Rules made there under and the charges on account of GST on these goods are correct under the provision of that Act or the rules made there under.

**1.6 Octroi Duty and / or other local taxes:** Contract Prices are all inclusive hence no reimbursement over and above the contract price(s) shall be allowed to seller towards payment of local taxes (such as levy of town duty, Octroi Duty, Terminal Tax and other levies of local bodies etc).

**1.7 Limitation of Liability:** The provisions of limitation of liability between Buyer and Seller as given in the GTC shall be applicable here.

**1.8 Resolution of disputes:** The provisions of DISPUTE RESOLUTION BETWEEN BUYER AND SELLER as given in the GTC shall be applicable here.

**1.9 Liquidated Damages:** If the Seller fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% per week or part of the week of delayed period as pre-estimated damages not exceeding 10% of the contract value without any controversy/dispute of any sort whatsoever. In case, Service Level Agreement (SLA) is applicable the same shall be applicable for the Contract.

#### 1.10 Financial Certificate:

**1.10.1** The expenditure involved for this purpose has received the Sanction of the competent financial authority.

**1.10.2** The funds are available under the proper head in the sanction budget allotment for the concern financial year.

**1.10.3** I have been fully authorized by the department to sign the supply order or incur the liability of the Goods being ordered.

**1.11** The bidder should submit a self declaration to the effect in bidder's official letter head that their agency have not been black listed by any Agency whatsoever till date.

### 2. Additional Terms and conditions

**2.1** Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

---

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.

---