

Organisation Details	Buyer Details
Type: Central PSU Ministry: Ministry of Coal Department: Materials Management Organisation Name: Bharat Coking Coal Limited Office Zone: Bcl Dhanbad	Name: Ashok Kumar Pandey Designation: Senior Manager(MM) Contact: - No.: - Email ID: ak.pandey4576@nic.in GSTIN: N MM Department, level 3, Commercial Building, Koyla Bhawan, Address: Koyla Nagar, DHANBAD, JHARKHAND-826005, India

Financial Approval Detail	Paying Authority Details
IFD Concurrence: Yes Designation of Administrative Approval: D(T)PP Designation of Financial Approval: D(T)PP	Name: BIKRAM GHOSH Designation: HOD Purchase Finance Email ID: hodpurfin.bcci@nic.in GSTIN: N MM Department, level 3, Commercial Building, Koyla Bhawan, Address: Koyla Nagar, DHANBAD, JHARKHAND-826005, India

Seller Details	
Company Name:	ELITE TRADING CO
Contact No.:	09811386789
Email ID:	Elitetradingcompany1000@gmail.com
Address:	H-9/A,SHAKARPUR,NEAR SANJAY PARK,SHAKARPUR NEAR SANJAY PARK, East Delhi, DELHI-110092, -
MSME verified:	No
GSTIN:	07HNBPS1312E1ZB

Product Details							
#	Item Description	Category Name	Model	HSN Code	Ordered Quantity	Unit	Price (Inclusive of all Duties and Taxes in INR)
1	BISTOS CTG Machine (Cardio Tocography) -Fetal Monitor Trolley mounted 24 hrs storage memory	CTG Machine (Cardio Tocography) -Fetal Monitor	BT 350	HSN not specified by seller	1	pieces	64,889
Total Order Value (in INR)							64,889

Consignee Detail						
S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Name: Anand Kumar Designation: - Email ID: anand.kumar@coalindia.in Contact: 0326-2206167-234 GSTIN: 20AAACB7934MFZB Address: Chief Of Medical Services, Central Hospital , PO- Jagjivan Nagar., DHANBAD, JHARKHAND-826003, India	BISTOS CTG Machine (Cardio Tocography) -Fetal Monitor Trolley mounted 24 hrs storage memory	-	1	09-Sep-2020	09-Oct-2020

Product Specification for BISTOS CTG Machine (Cardio Tocography) -Fetal Monitor Trolley mounted 24 hrs storage

memory

Specification	Sub-Spec	Value
GENERAL FEATURES	Product description	CTG Machine (Cardio Tocography)-Fetal Monitor
	Type	Trolley mounted
	Clinical purpose	For recording and analyzing the Fetal Heart Rate (FHR) and contractions of uterus during antepartum and intrapartum period
PRODUCT & FUNCTIONAL REQUIREMENTS	Advanced digital signal process	Yes
	Toco and Maternally sensed fetal movements are detected both manually	Yes
	Facility of twin monitoring	Yes
	Display of FHR upto twins FHR1, FHR2 and Toco	Yes
	Heart rate offset mode to visually distinguish between the heart rates of twins with ability to offset the secondary FHR by atleast 20bpm	Yes
	Capable to simultaneously provide integrated fetal heart rate (FHR) and Uterine activity	Yes
	Blinking correspondence to each beat	Yes
	In-built / Separate Acoustic Stimulator with a separate marker on the graph for acoustic stimulators	Yes
	Ultrasound transducer for measuring fetal heart rate with auto correlation	Yes
	Pulse Doppler transducer frequency (MHz)	=< 1.5
	Pulse Doppler with 7 to 9 elements/crystal design for better sensitivity	6 to 9
	Transducer should be waterproof	Yes
	Fetal Heart Rate measurements Range (BPM)	30 to 240
	FHR measurement tolerance (BPM)	+/-1
	Intensity of ultrasound transducer(mW/Sq Cm)	10
	User adjustable audio visual alarm for fetal heart rate (FHR)	Yes
	High alarm range (bpm)	160 to 180
	Low alarm range (bpm)	90 to 120
	Auto timer for NST tracing	Yes
	External Toco transducer unit, sealed waterproof and designed with guard ring to reduce maternal respiration artifact	Yes
	Strain gauge sensor element for measuring uterine tone contraction	Yes
	Toco Transducer measurement range, (Units)	0 to 100
	Resolution for transducer measurement, (Units)	+/-1
	Recorder for Fetal Heart Rate and Uterine contractions	Yes
	Computerized trace interpretation should be provided for antepartum monitoring of high risk pregnancies	Yes
	Digital display or waveform of Tabular/Graphical data of the patient containing the details viz Patient Name, ID, date and time for minium of 24 hours	Yes
	Suitable trolley with castor wheels, cover and locking facility with provision for parking the transducer etc	NA
	Size (LxWxH) (mmxmmxmm)	326X276X96
Weight (Kg)	2.8	
Power supply	220+/-10% V, 50/60 Hz, AC Single phase	

ELECTRICAL REQUIREMENTS	Input power (Watts)	80
	In-built rechargeable battery	Yes
	Type of battery	Li-ion
	Battery back up (hrs)	2
	Auto power off when machine is not in use for more than a preset time (minutes)	Yes
USER INTERFACE	Colour screen monitor type	LED
	Flat panel display, tiltable and foldable	Yes
	Touch screen display	No
	Display monitor size (inch)	7-8
	Built-in Printer to print patient data & Test Results	Yes
	Type of printer	High resolution Thermal printer
	Selectable Print speed for real time traces (Cm/minute)	1,2,3
	Print media	Z folded Thermo sensitive paper
	Print media sheet size (mmxmm)	150X90
	Patient data storage and trend graph for 48 to 72 hrs with fast printing facility of stored data	Yes
	Fast print speed for printing stored data (Cm/minute)	3
	FHR Chart printing width (mm)	150
	Toco Chart printing width (mm)	150
	Memory for storage of records (hrs)	24
	Machine should be compatible with wireless connection system for central monitoring	Yes
Machine should be compatible with wired connection system for central monitoring	Yes	
LIST OF ACCESSORIES	FHR probe (No)	2
	TOCO probe (No)	1
	Event Marker (No)	1
	Fetal Simulator (No)	1
	Probe belt (No)	3
	Ultrasound gel (No)	10
	Power Cable (No)	1
	Compatible size of z folded thermal sensitive continuous paper in packets of 150 sheets each (Number of Packets)	20
PACKING MODE	The product is packed individually in a box with all its accessories in such a way that there will be no damage during transportation	Yes
ENVIRONMENTAL CONSIDERATIONS	Storage conditions	Temperature -10 deg C to 40 deg C, RH not more than 85%
	Work or usage environment	Temperature 5 deg C to 40 deg C, RH not more than 85%
CERTIFICATIONS & REPORTS	Submission of Test Report from Central Govt/NABL/ILAC accredited Lab to prove the conformity to declared specification to the buyer on demand	YES
	Product certification	EU-CE
	Four digit number of notified body If product is EU-CE certified	4122
	Certificate No	4122

	Certificate date	14/8/18
	Certificate issuing authority	OTABU CERTIFICATION
	Certification, performance and safety standards specific to the device	EN60601-1,EN60601-1-2,EN60601-2-37
	Submission of all the certifications and test reports to the buyer on demand	Yes
INSTALLATION & TRAINING	Supplier to perform installation, safety and operation checks before handover	Yes
	Installation of machine required and to provide onsite service to end user	Yes
	Training of users in operation and basic maintenance shall be provided	Yes
WARRANTY & MAINTENANCE	Warranty (Yrs)	3
	User technical and maintenance manual detailing complete maintaining schedule with routine maintenance should be provided	Yes
	Contact details of manufacturer, supplier and local service agent to be provided	Yes

Corrigendum

1. **Extended Upto** : 2020-08-08 11:00:00

Terms and Conditions

1. General Terms and Conditions

1.1 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) as available on the GeM portal (unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

1.2 Terms of delivery: Free Delivery at Site including loading/unloading. In respect of items requiring installation and / or commissioning and other services in the scope of supply (as indicated in respective product category specification / STC / ATC), and the cost of the same is also included in the Contract price.

1.3 Delivery period: The Delivery Period/Time shall be essence of the Contract and delivery must be completed not later than such date(s). Any modification thereto shall be mutually agreed and incorporated in the Contract as per the provisions of the GTC.

1.4 Performance Security: If the Seller fails or neglects to observe or perform any of his obligations under the contract it shall be lawful for the Buyer to forfeit either in whole or in part, the Performance Security furnished by the Seller.

1.5 Taxes and Duties: Contract Prices are all inclusive i.e. including all taxes, duties, local levies / transportation / loading-unloading charges etc. Break up of GST shall be indicated by the Seller while raising invoice / bill on GeM. While submitting the bill / invoice Seller shall undertake that the Goods and Services Tax (GST) charged on this bill is not more than what is payable under the provision on the relevant Act or the Rules made there under and that the Goods on which GST has been charged have not been exempted under the GST Act or the Rules made there under and the charges on account of GST on these goods are correct under the provision of that Act or the rules made there under.

1.6 Octroi Duty and / or other local taxes: Contract Prices are all inclusive hence no reimbursement over and above the contract price(s) shall be allowed to seller towards payment of local taxes (such as levy of town duty, Octroi Duty, Terminal Tax and other levies of local bodies etc).

1.7 Limitation of Liability: The provisions of limitation of liability between Buyer and Seller as given in the GTC shall be applicable here.

1.8 Resolution of disputes: The provisions of DISPUTE RESOLUTION BETWEEN BUYER AND SELLER as given in the GTC shall be applicable here.

1.9 Liquidated Damages: If the Seller fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% per week or part of the week of delayed period as pre-estimated damages not exceeding 10% of the contract value without any controversy/dispute of any sort whatsoever. In case, Service Level Agreement (SLA) is applicable the same shall be applicable for the Contract.

1.10 Financial Certificate:

1.10.1 The expenditure involved for this purpose has received the Sanction of the competent financial authority.

1.10.2 The funds are available under the proper head in the sanction budget allotment for the concern financial year.

1.10.3 I have been fully authorized by the department to sign the supply order or incur the liability of the Goods being ordered.

1.11 The bidder should submit a self declaration to the effect in bidder's official letter head that their agency have not been black listed by any Agency whatsoever till date.

2. Additional Terms and conditions

2.1 Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.
